

Serial No. C 128592

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

Conmac Alloys.
Oldwell Ind est.
Irvine.
KALL SAR
LOCATION

Machine Details

Cust ref 1N
Make Bilanciai
Model D0700
Ser No 25070428
Capacity 50 1000 / 4
e/d 120 / 4
Working cap 44 000 g

Test Weights

Set No.	Cert No.	Date Cal
P01-18	7213159	NOV 25

Cert Interval	12M.
Next Cert	Feb 27.

Tolerance +/-	
0-10000	20
10000-40000	40
40000-50000	60

Comments
FLT - 5060.
WAGON - 21060

2	4	6	8
1	3	5	7

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
2000	2000	0	2000	0
6000	6000	0	6000	0
12000	12000	0	12000	0
18000	18000	0	18000	0
21060	21060	0	21060	0
31060	31060	0	31060	0
39060	39060	0	39060	0
44120	44120	0	44120	0

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind
44120	44120	44120	44120

Repeatability

Load	Ind	Ind	Ind

Eccentric Load 7060

Reading	Dev	Reading	Dev
1 7060	0	2 7060	0
3 7060	0	4 7060	0
5 7060	0	6 7060	0
7 7060	0	8 7060	0

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature Name Ian D. L. Date 20-2-26.

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



Members of the
UK Weighing Federation
Membership No. 1070

CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

Lowmac Alloy.
Oldhall West ind Est
Irvine.
K11 SAR.
LOCATION

Machine Details

Cust ref OUT
Make Bilanciai
Model DD700
Ser No 25190568
Capacity 50 / 000 / kg
e/d 1 20 / kg
Working cap 44000 kg

Test Weights

Set No.	Cert No.	Date Cal
PB1-18	TR13159	Nov 25

Cert Interval	Next Cert
12M.	Feb 27

Tolerance +/-	kg
0-10000	20
10000-40000	40
40000-50000	60

Comments

FCI - 3060.
WALTON - 21060.

7	5	3	1
8	6	4	2

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
9000	2000	0	2000	0
6000	6000	0	6000	0
17000	12000	0	12000	0
18000	18000	0	18000	0
21060	21060	0	21060	0
31060	31060	0	31060	0
39060	39060	0	39060	0
44120	44140	+20	44140	+20

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind
44120	44140	44140	44140

Repeatability

Load	Ind	Ind	Ind

Eccentric Load

Reading	Dev	Reading	Dev
1 7080	+20	2 7080	+20
3 7080	+20	4 7060	0
5 7060	0	6 7060	0
7 7060	0	8 7080	+20

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature Name IAN DICKSON Date 24.2.26